

## **CD163**

## IN VITRO DIAGNOSTIC DATASHEET

INTENDED USE: IN VITRO DIAGNOSTIC USE

This product is intended for qualitative immunohistochemistry with normal and neoplastic formalin-fixed, paraffin-embedded tissue sections, to be viewed by light microscopy.

**DESCRIPTION:** CD163 has been identified as an acute phase-regulated transmembrane protein whose function is to mediate the endocytosis of haptoglobin-hemoglobin complexes. This receptor is expressed on the surface of monocytes with low expression and on tissue macrophages, histiocytes with high expression. Staining with anti-CD163 has been helpful to distinguish synovial macrophages from synovial intimal fibroblasts in rheumatoid arthritis, where specificity for macrophages was found to be superior to that of CD68.

CATALOG NO: PL1103 PL1103-R7 7 ML RTU 70 TEST

PL1103-R1 1 ML RTU 10 TEST

STAINING PATTERN: Cytoplasmic and Membraneous PL1103-1 1 ML 1/100 1000 TEST

PL1103-0,1 0,1 ML 1/100 100 TEST

POSITIVE CONTROL: Liver

VOLUME: 7 ml Ready to Use (7 ml of antibody prediluted in 0.05mol/L Tris-HCl, pH 7.6 containing

stabilizing protein and 0.015mol/L sodium azide.)

**HOST:** Mouse

CLONE: ZM29

ANTIBODY CONCENTRATION: Not known

**SPECIES REACTIVITY:** Human. Others-not known.

**EPITOPE**: Not determined.

MICROBIOLOGICAL STATE: This product is not sterile.

**PRETREATMENT:** Staining of formalin-fixed tissue sections requires treating the tissue sections in boiling 10mM citrate buffer, pH 6.0 for 10-20 minutes followed by cooling at room temperature for 20 min.

**PRIMARY ANTIBODY INCUBATION TIME:** 30 minutes at Room Temperature

STAINING TIPS: If the staining is too light, use lower dilution or longer time. If the staining is

too strong, check pretreatment, use higher dilution or shorter time.

**STORAGE AND STABILITY:** This product contains sodium azide and is stable for 24 months when stored at 2-8°C. Do not use after expiration date indicated on label of the product. If reagent

is not stored as recommended, performance must be validated by the user.

**TROUBLESHOOTING:** Please contact Patolab Technical Support by e-mail ( patolab@patolab.com.tr ).



