

# IN VITRO DIAGNOSTIC DATASHEET

# PatoMark Black Tissue marking dye

**INTENDED USE :** For In Vitro Diagnostic Use

Catalog No	<u>Volume</u>
Black-50	50 ml
Black-500	500 ml
Black-1000	1000 ml

#### DESCRIPTION

Dyes for identification and orientation of excised surgical specimens formulated to define margins without bleeding, changing colour or fading and without staining fixatives and solvents used for tissue processing.

#### **STORAGE & SHELF LIFE**

Store at room temperature. Stable for minimum 3 years.

## **MICROBIOLOGICAL STATE**

Product(s) not sterile.

# MATERIALS REQUIRED BUT NOT PROVIDED

N/A

### **SPECIMEN & REAGENT PREPARATION**

Refer to Procedure.

### **PROCEDURE :**

1) Shake the bottle before use.

2) For fresh specimens (not fixed yet), apply the dye on the tissue, directly from the bottle or with the help of a brush. Let Dry for 2-3 minutes and Fix in Formalin; For fixed specimens, remove the excess of fixative from tissue surface and apply the dye like on fresh specimen;

3) The dyes bind to the tissue surface at room temperature (15-30°C) in 2-3 minutes;

4) At the end of the use, close the bottle to avoid evaporation.

### **Troubleshooting**

Please contact Patolab Technical Support by e-mail (patolab@patolab.com.tr).



IVD