

CD10

IN VITRO DIAGNOSTIC DATASHEET

INTENDED USE : IN VITRO DIAGNOSTIC USE

This product is intended for qualitative immunohistochemistry with normal and neoplastic formalin-fixed, paraffin-embedded tissue sections, to be viewed by light microscopy.

DESCRIPTION : CD10, also known as Common Acute Lymphocytic Leukemia Antigen (CALLA), is a cell surface enzyme with neutral metalloendopeptidase activity which inactivates a variety of biologically active peptides. CD10 is expressed on the cells of lymphoblastic, Burkitt's, and follicular germinal center lymphomas, and on cells from patients with chronic myelocytic leukemia (CML). It is also expressed on the surface of normal early lymphoid progenitor cells, immature B cells within adult bone marrow and germinal center B cells within lymphoid tissue.

CATALOG NO : PL728	PL728-R7 PL728-R1	7 ML RTU 70 TEST 1 ML RTU 10 TEST
STAINING PATTERN : Cell Membrane	PL728-1 PL728-0,1	1 ML 1/100 1000 TEST 0,1 ML 1/100 100 TEST
POSITIVE CONTROL : Small intestine, tonsil		0,2 1, 200 200 7207
VOLUME : 7 ml Ready to Use (7 ml of antibody prediluted in 0.05mol/L Tris-HCl, pH 7.6 containing stabilizing protein and 0.015mol/L sodium azide.)		
HOST: Mouse		
CLONE : 56C6		
ANTIBODY CONCENTRATION : Not known		
SPECIES REACTIVITY: Human. Others not tested.		
EPITOPE : Extracellular		
MICROBIOLOGICAL STATE : This product is not sterile.		
PRETREATMENT : Staining of formalin-fixed tissue sections requires treating the tissue sections in boiling 1mM EDTA, pH 8.0, for 10-20 min followed by cooling at RT for 20 min.		
PRIMARY ANTIBODY INCUBATION TIME : 60 minutes at Room Temperature		
STAINING TIPS : If the staining is too light, use lower dilution or longer time. If the staining is too strong, check pretreatment, use higher dilution or shorter time.		
STORAGE AND STABILITY : This product contains sodium azide and is stable for 24 months when stored at 2-8°C. Do not use after expiration date indicated on label of the product. If reagent is not stored as recommended, performance must be validated by the user.		
TROUBLESHOOTING : Please contact Patolab Technical Support by e-mail (patolab@patolab.com.tr).		

PatoLab Biyomedikal Cihazlar ve Sarf Malzemeler San. ve Tic. Ltd. Şti. patolab@patolab.com.tr www.patolab.com.tr

